

# 8401

## IEPE Simulator



## TECHNICAL SPECIFICATIONS

### Performance

<b>Input Voltage</b>	±5V
<b>Output Impedance</b>	1 x 108Ω to 1 x 109Ω
<b>Accuracy</b>	1.00 (+0, +2%)
<b>Output Bias</b>	11V ±3V DC
<b>Supply Voltage</b>	+18 to 28V DC
<b>Supply Current</b>	2 to 20mA
<b>Amplitude Linearity</b>	<2%
<b>Noise</b>	<40μV
<b>Size (mm)</b>	60 x 26 x 60mm
<b>(Inches)</b>	2.36 x 1.02 x 2.36"

DJB INSTRUMENTS UK LIMITED

e: sales@djbinstruments.com

t: +44 (0) 1638 712288

w: djbinstruments.com



SPECIFICATIONS MAY BE SUBJECT TO CHANGE WITHOUT PRIOR NOTICE.

ALL DEVICES CALIBRATED IN ACCORDANCE WITH BS ISO 16063-21:2003. ALL DEVICE CALIBRATIONS ARE UKAS TRACEABLE.